



CERTIFICATE OF ANALYSIS

PRODUCT NAME : DIAZO REAGENT A

Date of analysis : 01/10/19

Mfd. : Oct. 2019

Exp. : Sept. 2024

Product code : 14134

Batch No : MCR-19618

S. R	Test	Specification	Observation
1	Description	A clear colourless liquid.	A clear colourless liquid.
2	Suitability test	Passes test	Passes test

REMARK: THE PRODUCT COMPLIES WITH THE ABOVE SPECIFICATIONS

[Tested By]

[QC Incharge]

HEAD-QUALITY ASSUARANCE
(G.G.Salian)

Note: Computer generated image requires no sign.

MOLYCHEM Office: 78/80 Babu Genu Road, Souri Bldg., 2nd Floor, Mumbai – 400 002 (INDIA)
Tel: +91 22 4066 8111, 2205 7772 . Fax: + 91 22 2203 4442. Email : info@molychem.net
Works: Plot No. B/9, Badlapur M.I.D.C.,OPP.Shobhavat Textile Co.,
Kulgaon, Badlapur (E), Dist.-Thane-421 503