



CERTIFICATE OF ANALYSIS

PRODUCT NAME : D.P.X. MOUNTANT ARFOR MICROSCOPY

CAS No : 130- 12- 2

Date of analysis : 31/10/19

Mfd. : Oct. 2019

Product code : 23140

Batch No : MCR-19889

Exp. : Sept. 2024

S. R	Test	Specification	Observation
1	Description	A clear colourless viscous liquid	A clear colourless viscous liquid
2	Refractive index (n)²⁰_D	1.515 - 1.525	1.522
3	Acidity	0.05 ml N/1% max.	<0.05 ml N/1%

REMARK: THE PRODUCT COMPLIES WITH THE ABOVE SPECIFICATIONS

[Tested By]

[QC Incharge]

**HEAD-QUALITY ASSUARANCE
(G.G.Salian)**

Note: Computer generated image requires no sign.

**MOLYCHEM Office: 78/80 Babu Genu Road, Souri Bldg., 2nd Floor, Mumbai – 400 002 (INDIA)
Tel: +91 22 4066 8111, 2205 7772 . Fax: + 91 22 2203 4442. Email : info@molychem.net
Works: Plot No. B/9, Badlapur M.I.D.C.,OPP.Shobhavat Textile Co.,
Kulgaon, Badlapur (E), Dist.-Thane-421 503**