



## **CERTIFICATE OF ANALYSIS**

**PRODUCT NAME : KARL FISCHER REAGENT PYRIDINE FREE**

**Date of analysis : 12/09/21**  
**Mfd. : Sept. 2021**  
**Exp. : Aug. 2024**

**Product code : 31850**  
**Batch No. : MCM-4323**

<b>S.R</b>	<b>Test</b>	<b>Specification</b>	<b>Observation</b>
1	<b>Description</b>	Dark reddish brown solution.	Dark reddish brown solution.
2	<b>Effective Factor</b>	NLT 5 mg H <sub>2</sub> O/ml	7.12 mg H <sub>2</sub> O/ml
3	<b>Dilution test ( 5ml subst. + 5 ml water)</b>	The solution is clear and yellow coloured	The solution is clear and yellow coloured
4	<b>Dilution test ( 5ml subst. + 20 ml water)</b>	The solution is clear and yellow coloured	The solution is clear and yellow coloured

**REMARK:** THE PRODUCT COMPLIES WITH THE ABOVE SPECIFICATIONS

**[Tested By]**

**[QC Incharge]**

**HEAD-QUALITY ASSUARANCE**  
**(G.G.Salian)**

**Note:** Computer generated image requires no sign.