



CERTIFICATE OF ANALYSIS

PRODUCT NAME : WATER FOR HPLC & SPECTROSCOPY
(Filtered Through 0.2 μ Filter)

CAS No : 7732-18-5
Mol. Formula : H₂O
Mol. WT : 18
Date of analysis : 01/06/22

Product code : 32230
Batch No : MCM-4790
Mfd. : June 2022
Exp. : May 2025

S. R	Test	Specification	Observation
1	Description	Clear colourless liquid	Clear colourless liquid
2	pH	5 - 7	6.86
3	Residue on evaporation	0.0003% max	0.0002%
4	Transmission (1 cm cell vs Air)	200nm - 100% min 250nm - 100% min	107.9% 111.5%

REMARK: THE PRODUCT COMPLIES WITH THE ABOVE SPECIFICATIONS

[Tested By]

[QC Incharge]

HEAD-QUALITY ASSUARANCE
(G.G.Salian)

Note: Computer generated image requires no sign

MOLYCHEM Office: 78/80 Babu Genu Road, Souri Bldg., 2nd Floor, Mumbai – 400 002 (INDIA)
Tel: +91 22 4066 8111, 2205 7772 . Fax: + 91 22 2203 4442. Email : info@molychem.net
Works: Plot No. B/9, Badlapur M.I.D.C.,OPP.Shobhavat Textile Co.,
Kulgaon, Badlapur (E), Dist.-Thane-421 503