



CERTIFICATE OF ANALYSIS

PRODUCT NAME : AMMONIA BUFFER SOLUTION

Date of analysis : 16/06/22
Mfd. : June 2022
Exp. : May 2025

Product code : 11770
Batch No. : MCM-4822

S. R	Test	Specification	Observation
1	Description	Clear colorless solution	Clear colorless solution
2	Test as buffer	Passes test	Passes test
3	Excess Magnesium Sulphate	0.005% max	<0.005%

REMARK: THE PRODUCT COMPLIES WITH THE ABOVE SPECIFICATIONS

[Tested By]

[Q.C Incharge]

HEAD-QUALITY ASSUARANCE
(G.G.Salian)

Note: Computer generated image requires no sign.

MOLYCHEM Office: 78/80 Babu Genu Road, Souri Bldg., 2nd Floor, Mumbai – 400 002 (INDIA)
Tel: +91 22 4066 8111, 2205 7772 . Fax: + 91 22 2203 4442. Email : info@molychem.net
Works: Plot No. B/9, Badlapur M.I.D.C.,OPP.Shobhavat Textile Co.,
Kulgaon, Badlapur (E), Dist.-Thane-421 503