



CERTIFICATE OF ANALYSIS

PRODUCT NAME : GOWER'S REAGENT

For Erythrocyte count
(RBC diluting fluid)

Date of analysis : 22/09/22
Mfd. : Sept. 2022
Exp. : Aug. 2025

Product code : 31807
Batch No. : MCM-5043

S. R	Test	Specification	Observation
1	Description	A clear colourless solution.	A clear colourless solution.
2	Density (at 25°C)	1.037 – 1.046 g	1.039
3	Acetic acid content	about 1.12 %	1.12%
4	Suitability for use as stain	Passes test	Passes test

REMARK: THE PRODUCT COMPLIES WITH THE ABOVE SPECIFICATIONS

[Tested By]

[Q.C Incharge]

HEAD-QUALITY ASSUARANCE
(G.G.Salian)

Note: Computer generated image requires no sign.

MOLYCHEM Office: 78/80 Babu Genu Road, Souri Bldg., 2nd Floor, Mumbai – 400 002 (INDIA)
Tel: +91 22 4066 8111, 2205 7772 . Fax: + 91 22 2203 4442. Email : info@molychem.net
Works: Plot No. B/9, Badlapur M.I.D.C.,OPP.Shobhavat Textile Co.,
Kulgaon, Badlapur (E), Dist.-Thane-421 503