



CERTIFICATE OF ANALYSIS

PRODUCT NAME : D.P.X. MOUNTANT AR FOR MICROSCOPY

CAS No : 130- 12- 2

Date of analysis : 10/01/23

Mfd. : Jan. 2023

Product code : 23140

Batch No : MCM-5298

Exp. : Dec. 2027

S. R	Test	Specification	Observation
1	Description	A clear colourless viscous liquid	A clear colourless viscous liquid
2	Refractive index (n) ²⁰ D	1.515 - 1.525	1.516
3	Acidity	0.05 ml N/1% max.	<0.05 ml N/1%

REMARK: THE PRODUCT COMPLIES WITH THE ABOVE SPECIFICATIONS

[Tested By]

**CHEMIST
(Sonu)**

[QC Incharge]

**HEAD-QUALITY ASSUARANCE
(G.G.Salian)**

Note: Computer generated image requires no sign.