



## **CERTIFICATE OF ANALYSIS**

**PRODUCT NAME : BUFFER SOLUTION pH 4.0 (PHTHALATE)**

**Date of Analysis : 09/03/23**  
**Mfd. : March 2023**  
**Exp. : Feb. 2026**

**Product code : 31390A**  
**Batch No. : MCM-5432**

| S.R | Test                               | Specification              | Observation                |
|-----|------------------------------------|----------------------------|----------------------------|
| 1   | Description                        | Clear colourless solution. | Clear colourless solution. |
| 2   | pH of solution (20 <sup>0</sup> C) | 4.0 ± 0.02                 | 4.01                       |

**REMARK:** THE PRODUCT COMPLIES WITH THE ABOVE SPECIFICATIONS

**[Tested By]**

**CHEMIST**  
**(Sonu)**

**[QC Incharge]**

**HEAD-QUALITY ASSURANCE**  
**(G.G.Salian)**

**Note:** Computer generated image requires no sign.

**MOLYCHEM Office: 78/80 Babu Genu Road, Souri Bldg., 2<sup>nd</sup> Floor, Mumbai – 400 002 (INDIA)**  
**Tel: +91 22 4066 8111, 2205 7772 . Fax: + 91 22 2203 4442. Email : [info@molychem.net](mailto:info@molychem.net)**  
**Works: Plot No.B/9, Badlapur M.I.D.C.,OPP.Shobhavat Textile Co.,**  
**Kulgaon, Badlapur (E), Dist.-Thane-421 503**