

CERTIFICATE OF ANALYSIS

PRODUCT NAME : MAGNESIUM CHLORIDE (HEXAHYDRATE) AR

| | | | |
|-------------------------|--|---------------------|--------------|
| CAS No | : 7791-18-6 | Product code | : 23880 |
| Mol. Formula | : MgCl ₂ .6H ₂ O | Batch No. | : MCRT-30576 |
| Mol. WT | : 203.30 | Mfd. | : March 2024 |
| Date of analysis | : 09/03/24 | Exp. | : Feb. 2029 |

| S.R | Test | Specification | Observation |
|-----|-----------------------------------|--|----------------------|
| 1 | Description | Colourless deliquescent crystals or white crystalline powder | Colourless crystals. |
| 2 | Assay (ex Cl) | 98.0% min | 99.83% |
| 3 | Alcohol-Insoluble matter | passes test | passes test |
| 4 | Water-insoluble matter | 0.003% max | 0.002% |
| 5 | Acidity (HCl) | 0.001 % max | <0.001 % |
| 6 | Alkalinity (MgO) | 0.0005% max | <0.0005% |
| 7 | Nitrogen compounds (N) | 0.002% max | <0.002% |
| 8 | Phosphate (PO₄) | 0.0002% max | <0.0002% |
| 9 | Sulphate (SO₄) | 0.002% max | <0.002% |
| 10 | Arsenic (As) | 0.0001% max | <0.0001% |
| 11 | Barium (Ba) | 0.005% max | <0.005% |
| 12 | Calcium (Ca) | 0.002% max | <0.002% |
| 13 | Copper (Cu) | 0.0002% max | <0.0002% |
| 14 | Iron (Fe) | 0.0002% max | <0.0002% |
| 15 | Lead (Pb) | 0.0002% max | <0.0002% |
| 16 | Manganese (Mn) | 0.0005% max | <0.0005% |
| 17 | Potassium (K) | 0.005% max | <0.005% |
| 18 | Sodium (Na) | 0.005% max | <0.005% |
| 19 | Zinc (Zn) | 0.001% max | <0.001% |

REMARK: THE PRODUCT COMPLIES WITH THE ABOVE SPECIFICATIONS

[Tested By]

[QC Incharge]

**CHEMIST
(Aishwarya)**

**HEAD-QUALITY ASSURANCE
(G.G.Salian)**

Note: Computer generated image requires no sign

MOLYCHEM H.Q. : Maker Tower 'E', 41-A2, 4th Floor, Above Maker Arcade, Cuffe Parade, Mumbai 400 005 (INDIA)

Tel. : +91 22 4066 8111 • Email: info@molychem.net • Web: www.molychem.net

Reg. Office : 78/80 Babu Genu Road, Souri Bldg., 2nd Floor, Mumbai - 400 002 (INDIA)

Works : Unit I - Plot No. B/9, Unit II - Plot No. B/11, Badlapur M.I.D.C., Opp. Shobhavat Textile Co., Kulgaoon, Badlapur (E), Dist.-Thane-421 503.