



CERTIFICATE OF ANALYSIS

Date: 15/10/16

PRODUCT NAME : D.P.X. MOUNTANT ARFOR MICROSCOPY

CAS No : 130- 12- 2

Product code : 23140

Mfd. : Oct. 2016

Batch No : MCR-12475

Exp. : Sept. 2021

S. R	Test	Specification	Observation
1	Refractive index (n) ²⁰ _D	1.515 - 1.525	1.525
2	Acidity	0.05 ml N/1% max.	<0.05 ml N/1%

REMARK: THE PRODUCT COMPLIES WITH THE ABOVE SPECIFICATIONS

[Tested By]

[QC Incharge]

HEAD-QUALITY ASSUARANCE
(G.G.Salian)

Note: Computer generated image requires no sign.