

CERTIFICATE OF ANALYSIS

PRODUCT NAME : WATER FOR HPLC & SPECTROSCOPY
(Filtered Through 0.2 μ Filter)

CAS No	: 7732-18-5	Product code	: 32230
Mol. Formula	: H ₂ O	Batch No	: MCM-7224
Mol. WT	: 18	Mfd.	: April 2025
Date of analysis	: 04/04/25	Exp.	: March 2026

S.R	Test	Specification	Observation
1	Description	Clear colourless liquid.	Clear colourless liquid.
2	pH	5 - 7	6.24
3	Residue on evaporation	max 0.0003%	0.0002%
4	Conductivity	max. 3.0 μ s	1.81 μ s
5	Transmission (1 cm cell vs Air)		
	200 nm	min. 100%	107.78%
	250 nm	min. 100%	109.84%

REMARK: THE PRODUCT COMPLIES WITH THE ABOVE SPECIFICATIONS

[Tested By]



**CHEMIST
(Akash)**



[QC Incharge]



**HEAD-QUALITY ASSURANCE
(G.G.Salian)**

MOLYCHEM H.Q. : Maker Tower 'E', 41-A2, 4th Floor, Above Maker Arcade, Cuffe Parade, Mumbai 400 005 (INDIA)

Tel. : +91 22 4066 8111 • Email: info@molychem.net • Web: www.molychem.net

Reg. Office : 78/80 Babu Genu Road, Souri Bldg., 2nd Floor, Mumbai - 400 002 (INDIA)

Works : Unit I - Plot No. B/9, Unit II - Plot No. B/11, Badlapur M.I.D.C., Opp. Shobhavat Textile Co., Kulgaon, Badlapur (E), Dist.-Thane-421 503.