

CERTIFICATE OF ANALYSIS

PRODUCT NAME : WATER FOR HPLC & SPECTROSCOPY
(Filtered Through 0.2 μ Filter)

CAS No : 7732-18-5
Mol. Formula : H₂O
Mol. WT : 18
Date of analysis : 18/06/25

Product code : 32230
Batch No : MCM-7420
Mfd. : June 2025
Exp. : May 2026

S.R	Test	Specification	Observation
1	Description	Clear colourless liquid.	Clear colourless liquid.
2	pH	5 - 7	6.58
3	Residue on evaporation	max 0.0003%	0.0002%
4	Conductivity	max. 3.0 μ s	2.24 μ s
5	Transmission (1 cm cell vs Air)		
	200 nm	min. 100%	111.61%
	250 nm	min. 100%	112.52%

REMARK: THE PRODUCT COMPLIES WITH THE ABOVE SPECIFICATIONS

[Tested By]



CHEMIST
(Deepa)



[QC Incharge]



HEAD-QUALITY ASSURANCE
(G.G.Salian)

MOLYCHEM H.Q. : Maker Tower 'E', 41-A2, 4th Floor, Above Maker Arcade, Cuffe Parade, Mumbai 400 005 (INDIA)

Tel. : +91 22 4066 8111 • Email: info@molychem.net • Web: www.molychem.net

Reg. Office : 78/80 Babu Genu Road, Souri Bldg., 2nd Floor, Mumbai - 400 002 (INDIA)

Works : Unit I - Plot No. B/9, Unit II - Plot No. B/11, Badlapur M.I.D.C., Opp. Shobhavat Textile Co., Kulgaon, Badlapur (E), Dist.-Thane-421 503.