

CERTIFICATE OF ANALYSIS

PRODUCT NAME : WATER DISTILLED


CAS No : 7732-18-5
Mol. Formula : H₂O
Mol. WT : 18
Date of analysis : 25/06/25

Product code : 32229
Batch No : MCM-7437
Mfd. : June 2025
Exp. : May 2026

S.R	Test	Specification	Observation
1	Description	Clear colourless liquid	Clear colourless liquid
2	pH	5.5 – 7	6.73
3	Density	About 1.000	1.000
4	Chloride (Cl)	max.0.0001%	< 0.0001%
5	Substance reducing KMnO₄	max. 0.00004%	< 0.00004%
6	Conductivity	max. 3.0 μs	2.5 μs

REMARK: THE PRODUCT COMPLIES WITH THE ABOVE SPECIFICATIONS

[Tested By]



**CHEMIST
(Jyoti)**



[QC Incharge]



**HEAD-QUALITY ASSURANCE
(G.G.Salian)**

MOLYCHEM H.Q. : Maker Tower 'E', 41-A2, 4th Floor, Above Maker Arcade, Cuffe Parade, Mumbai 400 005 (INDIA)

Tel. : +91 22 4066 8111 • Email: info@molychem.net • Web: www.molychem.net

Reg. Office : 78/80 Babu Genu Road, Souri Bldg., 2nd Floor, Mumbai - 400 002 (INDIA)

Works : Unit I - Plot No. B/9, Unit II - Plot No. B/11, Badlapur M.I.D.C., Opp. Shobhavat Textile Co., Kulgaon, Badlapur (E), Dist.-Thane-421 503.