



CERTIFICATE OF ANALYSIS

Date: 15/03/17

PRODUCT NAME : ORANGE G INDICATOR FOR MICROSCOPY

CAS No : 1936-15-8
Mfd. : March 2017
Exp. : Feb. 2022

Product code : 32054
Batch No : MCR-13357

S. R	Test	Specification	Observation
1	Absorption Maxima (Water)	476 - 485 nm	477.70 nm
2	Specific absorptivity A1% / 1 cm l _{max} 0.0015 % Water	NLT 350	450.19
3	Loss on drying at 110°C	15% max	9.3%
4	Suitability for microscopy	Passes test	Passes test

REMARK: THE PRODUCT COMPLIES WITH THE ABOVE SPECIFICATIONS

[Tested By]

[QC Incharge]

HEAD-QUALITY ASSUARANCE
(G.G.Salian)

Note: Computer generated image requires no sign.

MOLYCHEM Office: 78/80 Babu Genu Road, Souri Bldg., 2nd Floor, Mumbai – 400 002 (INDIA)
Tel: +91 22 4066 8111, 2205 7772 . **Fax:** + 91 22 2203 4442. **Email :** info@molychem.net
Works: Plot No. B/9, Badlapur M.I.D.C., OPP.Shobhavat Textile Co.,
Kulgaon, Badlapur (E), Dist.-Thane-421 503