CERTIFICATE OF ANALYSIS

Date: 13/07/18

PRODUCT NAME : CHROMOTROPIC ACID DISODIUM SALT AR

CAS No : 5808-22-0
Mol. Formula : $C_{10}H_6Na_2O_8S_2.2H_2O$
Mol. WT : 400.29
Exp. : June 2023

Product code : 22645
Batch No : MCR-9187-03
Mfd. : July 2018

S. R	Test	Specification	Observation
1	**Description**	Brown coloured powder	Brown coloured powder
2	**Assay**	98.5% min.	98.58%
3	**Sensitivity to Nitrate**	Passes test	Passes test
4	**Sensitivity to Formaldehyde**	Passes test	Passes test
5	**Sulphated ash**	35 – 36%	35.2%
6	**Sulphate (SO_4)**	0.002% max	<0.002%
7	**Water (Karl fischer)**	8.5 – 9.5%	9.42%

REMARK: THE PRODUCT COMPLIES WITH THE ABOVE SPECIFICATIONS

[Tested By] **[QC Incharge]**

HEAD-QUALITY ASSUARANCE
(G.G.Salian)

Note: Computer generated image requires no sign

MOLYCHEM Office: 78/80 Babu Genu Road, Souri Bldg., 2nd Floor, Mumbai – 400 002 (INDIA)
Tel: +91 22 4066 8111, 2205 7772 . Fax: + 91 22 2203 4442. Email : info@molychem.net
Works: Plot No. B/9, Badlapur M.I.D.C.,OPP.Shobhavat Textile Co.,
Kulgaon, Badlapur (E), Dist.-Thane-421 503